


PULSE GENERATOR PERFORMANCE
CHECK

Model: AVI-V
S.N.: 127
Date: 27/9/77



12 VOLT/cm
10 nsec/cm.

- a) Output signal amplitude: 50 volts.
- b) Pulse width: 0-100 nsec.
- c) Rise time: 0.5 nsec.
- d) Fall time: ≤ 2.0 nsec.
- e) PRF: ≤ 20 KHz.
- f) Jitter, stability: OK
- g) Prime power: $+15 \pm 1$ volt.
 ≤ 250 mA.