


IMPULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5

S.N.: 130

Date: OCT 13 1977



1.0 NSEC/DIV.
6 VOLTS/DIV.

- a) Output signal amplitude:
 ≈ 20 VOLTS
- b) Pulse width: 0.5 NSEC
 (10% RISE TIME)
- c) Jitter, stability:
 OK
- d) Prime power:
 $+15 \pm 1$ V
 145 mA.