


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVI-V
S.N.: 133
Date: 16/12/77



10 nsec/cm.

- a) Output signal amplitude:
 ≈ 50 VOLTS.
- b) Pulse width:
 0-100 nsec.
- c) Rise time:
 0.5 nsec.
- d) Fall time:
 ≤ 2.0 nsec
- e) PRF:
 ≤ 20 kHz
- f) Jitter, stability:
 OK.
- g) Prime power:
 $\pm 15 \pm 1$ VOLT
 180 mA.