


PULSE GENERATOR PERFORMANCE
CHECK

Model: AVI-MP
S.N.: 143
Date: 20/2/78



5 nsec/div

- a) Output signal amplitude:
 > 40 VOLTS
- b) Pulse width: VARIABLE FROM
 1.0 TO 100 nSEC.
- c) Rise time: ≤ 0.5 nSEC.
- d) Fall time: ≤ 2.0 nSEC.
- e) PRF: OK.
- f) Jitter, stability: OK.
- g) Prime power:
 $+15 \pm 1$ VOLT
 @ 120 mA

[Handwritten signature]