


PULSE GENERATOR PERFORMANCE
CHECK

Model: AVI-V
S.N.: 144
Date: 20/2/78



- a) Output signal amplitude:
 > 50 VOLTS.
- b) Pulse width:
 $0-100$ NSEC.
- c) Rise time: ≤ 0.5 NSEC.
- d) Fall time: ≤ 2.0 NSEC.
- e) PRF: ≤ 20 KHZ
- f) Jitter, stability: OK
- g) Prime power:
 $\pm 15 \pm 1$ VOLT
 ≤ 200 mA