


IMPULSE GENERATOR PERFORMANCE

CHECK

Model: AVH
S.N.: 145
Date: 3/4/78



1 nsec/cm.

- a) Output signal amplitude
 ≥ 30 VOLTS
- b) Pulse width ≤ 1.0 NSEC
 (10% RISE TIME)
- c) Prime power
 15 ± 1 VOLT
 135 mA
- d) Jitter, stability

OK