


IMPULSE GENERATOR PERFORMANCE

CHECK

Model: AVH-MPT

S.N.: 149

Date: 8/5/78



1.0 nsec/div  
≈ 7.0 volts/div

a) Output signal amplitude

7.30 VOLTS

b) Pulse width

≤ 1.0 nsec  
(10% RISE TIME)

c) Prime power

+15 ± 1 VOLT  
< 150 mA

d) Jitter, stability

OK