


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVK-A-P

S.N.: 160

Date: SEPT 20 78



1.0 nsec/DIV
~ 8 VOLTS/DIV.

- a) Output signal amplitude
 $> 15 \text{ V}$
- b) Pulse width
 $4 - 100 \text{ nsec.}$
- c) Rise time:
 $< 150 \text{ nsec.}$
- d) Fall time:
 OK.
- e) PRF:
 OK
- f) Jitter, stability:
 OK
- f) Prime power
 $+15 \pm 1 \text{ V}$
 290 mA