


PULSE GENERATOR PERFORMANCE

CHECK

Model: *AVK-1A-N*  
S.N.: *161*  
Date: *SEPT 20 78*



*1.0 nsec/div*  
*≈ 8.0 volts/div.*

- a) Output signal amplitude  
*> 15 volts.*
- b) Pulse width:  
*4-100 nsec.*
- c) Rise time:  
*< 100 nsec.*
- d) Fall time:  
*OK*
- e) PRF:  
*OK*
- f) Jitter, stability:  
*OK.*
- g) Prime power  
*+15 ± 1 volt.*  
*240 mA*