


url
WPT

Pulse Generator Performance

Check

Model: *AVK-A-N*
S.N. *172*
Date: *DEC 5 1978*



1 NSEC/DIV
≈ 15 VOLTS/DIV

- a) Output signal amplitude
≤ 15 VOLTS
- b) Pulse width:
4 - 100 NSEC
- c) Rise time:
≤ 100 PSEC
(≈ 70 PSEC)
- d) Fall Time:
≤ 1.0 NSEC
- e) PRF:
OK
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1 VOLT
≈ 200 mA