


4 copies

IMPULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5

S.N.: 174

Date: DEC 29 78



a) Output signal amplitude:

7.20 VOLTS

b) Pulse width:

0.5 NSEC.
(10% RISE TIME)

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1 VOLTS

120 mA

1.0 NSEC / DIV.
7.0 VOLTS / DIV