


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *178*

Date: *FEB 8 79*



a) Output signal amplitude:

*> 30 VOLTS*

b) Pulse width:

*≤ 1.0 nSEC*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1 VOLTS*

*160 mA*

*1.0 nSEC/DIV*

*≈ 7.0 VOLTS/DIV*