


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MP*

S.N.: *179*

Date: *FEB 8 79*



*1.0 NSEC / DIV*  
*~ 7.5 VOLTS / DIV*

a) Output signal amplitude:

*> 30 VOLTS*

b) Pulse width:

*≤ 1.0 NSEC*  
*(10% rise)*

c) Jitter, stability:

*OK*

d) Prime power:

*+ 15 ± 1 VOLT*

*100 mA*

*ALC*