


Pulse Generator Performance

Check

Model: *AVK-A-12*

S.N.: *180*

Date: *FEB 8 79*



*1.0 NSEC/DIV
7.0 VOLTS/DIV
(MAX OUTPUT)*

- a) Output signal amplitude:
15 VOLTS
- b) Pulse width:
4-100 NSEC.
- c) Rise time:
< 100 PSEC.
- d) Fall Time:
< 1.0 NSEC.
- e) PRF:
*0 - 1 MHz FULL
MAN PW*
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1 VOLTS
195 mA.*