


IMPULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5

S.N.: 188

Date: FEB 22 79



1.0 nsec / DIV
≈ 6 VOLTS / DIV.

a) Output signal amplitude:

> 20 VOLTS

b) Pulse width:

0.5 nsec
(10% RISE TIME)

c) Jitter, stability:

OK.

d) Prime power:

+15 ± 1 V

160 mA