


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MP

S.N.: 198

Date: APRIL 16 '79



1 nsec/cm  
≈ 7 VOLTS/cm

a) Output signal amplitude:

≈ 30 VOLTS.

b) Pulse width:

≤ 1.0 nsec

(10% RISE TIME)

c) Jitter, stability:

OK

d) Prime power

+15 ± 1 VOLT

135 mA