


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V*

S.N.: *200*

Date: *APRIL 21 1979*



- a) Output signal amplitude:
> 50 VOLTS
- b) Pulse width:
0 - 100 NSEC.
- c) Rise time:
≤ 0.5 NSEC.
- d) Fall time:
≤ 2.0 NSEC.
- e) PRF:
0 - 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1 VOLT

10 NSEC/DIV
≈ 17 VOLTS/DIV

240 mA