


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVL-AV

S.N.: 210

Date: MAY 30 '79



≈ 70 VOLTS/DIV
10 NSEC/DIV.

- a) Output signal amplitude:
50-250 VOLTS
- b) Pulse width:
0-100 NSEC.
- c) Rise time:
≤ 5 NSEC
- d) Fall time:
≤ 5 NSEC.
- e) PRF:
0-5 KHz
- f) Jitter, stability:
OK.
- g) Prime power:
+15 ± 1 VOLT
500 mA (MAX)