


PULSE GENERATOR PERFORMANCE

CHECK

Model: *AVK-V-P*

S.N.: *212*

Date: *MAY 30 79*



1.0 nsec/div
2 1/2 13 volts/div.

- a) Output signal amplitude: *
- b) Pulse width:
- c) Rise time:
- d) Fall time:
- e) PRF:
- f) Jitter, stability:
- g) Prime power:

15 VOLTS

0 - 100 nsec

≤ 100 psec

(10% RISE TIME)

≤ 1.0 nsec.

0 - 50 kHz

OK.

+15 ± 1 VOLT

225 mA

* N.B. EXCEEDS 5 VOLT
57% 10 db.