


PULSE GENERATOR PERFORMANCE

CHECK

Model: AVK-V-N

S.N.: 213

Date: MAY 30 '79



1.0 NSEC/DIV.
≈ 13 VOLTS/DIV.

- a) Output signal amplitude: *
15 VOLTS
- b) Pulse width:
0-100 NSEC.
- c) Rise time:
≤ 100 PSEC
(10% RISE TIME)
- d) Fall time:
≤ 10 NSEC.
- e) PRF:
0-50 KHZ.
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1 VOLT
250 mA

[Signature]
* N.B. EXCEEDS 5 VOLT
SPEC BY 10 db.