


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-U-A*

S.N.: *220*

Date: *JULY 27 79*



*1.0 nSEC / DIV.
≈ 7.0 VOLTS / DIV.*

- a) Output signal amplitude:
6.8 VOLTS
- b) Pulse width:
0-50 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time: (*100% RISE*)
≤ 2.0 nSEC
- e) PRF:
0-50 kHz
- f) Jitter, stability:
6K
- g) Prime power:
*+15 ± 1.0 VOLT
275 mA*