


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT-N

S.N.: 248

Date: OCT 15 1979



1.0 nsec/div
1.2 volts/div

- a) Output signal amplitude:
 > 30 VOLTS
- b) Pulse width:
 ≤ 1.0 nsec
- c) Rise time: (10% RISE TIME)
 NA
- d) Fall time:
 NA
- e) PRF:
 0 TO 1.0 MHz
- f) Jitter, stability:
 OK
- g) Prime power:

+15 ± 1 VOLT

180 mA