


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-V-1-S

S.N.: 257

Date: DEC 14 79



1.0 nsec/DIV

20.0 VOLTS/DIV

200 KHz

- a) Output signal amplitude:
0 TO +30V
- b) Pulse width:
0 TO 4.0 nsec
- c) Rise time:
≤ 150 nsec
- d) Fall time:
≤ 150 nsec
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1 VOLT

400 mA