


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *263*

Date: *DEC 14 79*



100 nsec/DIV
2.6 VOLTS/DIV

- a) Output signal amplitude:
> 30 VOLTS
- b) Pulse width:
≤ 1.0 nsec
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+ 15 ± 1.0 VOLT
160 mA