


PULSE GENERATOR  
PERFORMANCE CHECK

Model: 1904-MPT

S.N.: 264

Date: DEC 14 79



1.0 nsec/div  
≈ 6 volts/div

- a) Output signal amplitude:  
    > 30 VOLTS
- b) Pulse width:  
    ≤ 1.0 nsec
- c) Rise time:  
    N/A.
- d) Fall time:  
    N/A.
- e) PRF:  
    0 TO 1.0 MHz
- f) Jitter, stability:  
    OK
- g) Prime power:

+15 ± 1 VOLT  
160 mA.

A large, stylized handwritten signature in black ink, consisting of several loops and a long horizontal stroke extending to the right.