


PULSE GENERATOR
PERFORMANCE CHECK

Model: AUP-AU-1-N

S.N.: 284

Date: JAN 30 80



20 VOLTS/DIV
1.0 NS/DIV

- a) Output signal amplitude:
 > -10 VOLTS (MIN)
- b) Pulse width:
 0 TO 4 NS
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 100 PSEC
- e) PRF:
 0 TO 10 MIPR
- f) Jitter, stability:
 OK
- g) Prime power:
 +15 ± 1.0 VOLT

275 mA