


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *289*

Date: *FEB 22 80*



a) Output signal amplitude:

*> 30 VOLTS*

b) Pulse width:

*≤ 1.0 nSEC*

*(10% RISE)*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 VOLT*

*175 mA*