


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-2-P-D*

S.N.: *304*

Date: *MAR 26 80*



*10 VOLTS / DIV  
1.0 NSEC / DIV*

- a) Output signal amplitude:  
*0 TO +15 VOLT*
- b) Pulse width:  
*0.2 TO 2.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 130 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+24 ± 1.0 VOLT  
0.6 AMP (MAX)*

*[Signature]*