


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-1-P

S.N.: 317

Date: APRIL 11 80



1.0 nsec/div
10 volts/div

- a) Output signal amplitude:
0 TO +10 V
- b) Pulse width:
0.2 TO 4.0 nsec
- c) Rise time:
≤ 100 psec (10% rise)
- d) Fall time:
≤ 130 psec
- e) PRF:
0 - 1.0 MHz
- f) Jitter, stability:
OK.
- g) Prime power:

+15 ± 1 VOLT

250 mA