


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT

S.N.: 324

Date: APRIL 28 80



a) Output signal amplitude:

+34 VOLTS

b) Pulse width:

$\leq 1.0 \text{ nsec}$ (10% rise)

c) Jitter, stability:

OK

d) Prime power:

+15 \pm 1.0 VOLT

200 mA

10 VOLTS/cm
1.0 nsec/cm

OVERLOADED SIGNALING SCOPE
TO REMOVE USE 50 dB
RATHER THAN 40 dB ATTENUATION.