


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *325*

Date: *APRIL 28 80*



- a) Output signal amplitude:
+ 37 VOLTS
- b) Pulse width:
≤ 1.0 NSEC (10% RISE)
- c) Jitter, stability:
OK
- d) Prime power:
+15 ± 1.0 VOLT
175 mA.

10 VOLTS/CM
1.0 NSEC/CM

*OVERLOADED SAMPLING SCOPE
TO REMOVE, USE 50 dB
ATTENUATOR THAN 40 dB ATTENUATION.*