


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *374*

Date: *AUG 27 80*



- a) Output signal amplitude:
> 30 VOLTS
- b) Pulse width:
≤ 1.0 nSEC
- c) Jitter, stability:
*(10% RISE)
OK*
- d) Prime power:
*+ 15 ± 1.0 V
175 mA (UMAX)*

*1.0 nsec/cm
≈ 8 VOLTS/cm
1.0 nsec*