


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-1-P

S.N.: 388

Date: SEPT 25 80



10V / DIV  
100 nSEC / DIV  
400 KITE:

- a) Output signal amplitude:  
0 TO +10V
- b) Pulse width:  
0.2 TO 4.0 nSEC
- c) Rise time:  
≤ 100 pSEC
- d) Fall time:  
≤ 130 pSEC
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

+15 ± 1.0 VOLTS

240 mA max PA