


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *440*

Date: *JAN 28 '81*



a) Output signal amplitude:

$\geq +30 \text{ V}$

b) Pulse width:

$\leq 1.0 \text{ nsec}$

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0 \text{ Volt}$

150 mA (MAX)

32 volts/cm

2.0 nsec/cm

400 KHz