


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: PVH-MPT-2

S.N.: 441

Date: JAN 28 81



a) Output signal amplitude:

$\geq +60$  VOLT

b) Pulse width:

$\leq 2.0$  NSEC

c) Jitter, stability:

OK.

d) Prime power:

$+15 \pm 1.0$  V

150 mA (MAX)

32 VOLTS/cm

2.0 nsec/cm

400 KHz.