


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-MPT*

S.N.: *443*

Date: *JAN 28 '81*



a) Output signal amplitude:

*≥ +30 VOLTS*

b) Pulse width:

*≤ 1.0 NSEC*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 V*

*150 mA (MAX)*

*32 VOLTS/cm*

*2.0 NSEC/cm*

*400 KHz*