


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-MPT

S.N.: 444

Date: JAN 28 81



a) Output signal amplitude:
 $\geq +30$ VOLTS.

b) Pulse width:
 ≤ 1.0 NSEC.

c) Jitter, stability:
OK

d) Prime power:

$+15 \pm 1.0$ V

160 mA (MAX)

32 VOLTS/cm

2.0 NSEC/cm

400 KHZ