


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-1

S.N.: 445

Date: FEB 18 81



a) Output signal amplitude:

+ 11 VOLTS

b) Pulse width:

130 - 150 PSEC.

c) Jitter, stability:

OK

d) Prime power:

+ 15 ± 1.0 VOLT
150 mA (max)

10 VOLTS/cm.
500 PSEC/cm
400 KHZ.