


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-AV-1-N*

S.N.: *460*

Date: *FEB 27 81*



- a) Output signal amplitude:
0 TO -125 V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
< 1 NSEC
- d) Fall time:
< 2 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT

325 mA (MAX)

*10 NSEC/DIV
100 VOLTS/DIV
4 KHZ*

[Signature]