


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV14-S-1-N*

S.N.: *469*

Date: *MAR 12 81*



a) Output signal amplitude:

*0 TO -10V*

b) Pulse width:

*≈ 150 PSEC*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 1.0 VOLT*

*500 PSEC/DIV*

*10 VOLTS/DIV*

*400 KHZ*