


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-1

S.N.: 476

Date: MAR 31 81



a) Output signal amplitude:

$\geq +10$ VOLTS

b) Pulse width:

≤ 150 PSEC

c) Jitter, stability:

OK

d) Prime power:

$+15 \pm 1.0$ VOLT.

200 mA

10 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ