


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S-1-A-05

S.N.: 482

Date: APRIL 14 81



a) Output signal amplitude:

0 TO -10V

b) Pulse width:

≈ 150 PSEC


c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 VOLT

200 mA ~~max~~



A large, stylized handwritten signature in black ink, located in the bottom right corner of the page.