


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVK-AU-N-05*


S.N.: *498*

Date: *APRIL 20 81*



*10 VOLTS/DIV.  
10 NSEC/DIV  
20 KHZ*

- a) Output signal amplitude:  
*0 TO -15 V*
- b) Pulse width:  
*0 TO 50 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:



*10 VOLTS/DIV.  
1.0 NSEC/DIV  
20 KHZ.*

*+15 ± 1.0 VOLT*

*290 mA (MAX)*