


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: **AVH-S-1-05-P**

S.N.: **500**

Date: **APRIL 29 81**



**10 VOLTS / DIV
1.0 NSEC / DIV
400 KHZ**

a) Output signal amplitude:

+10 V

b) Pulse width:

≤ 150 PSEC


c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 WOLT

210 mA (MAX)



**10 VOLTS / DIV
100 PSEC / DIV
400 KHZ**

A large, stylized handwritten signature or set of initials in dark ink, located to the right of the second oscilloscope trace.