


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-HVA*

S.N.: *511*

Date: *APRIL 28 81*



a) Output signal amplitude:

*0 to +90 V*

b) Pulse width:

*≤ 1.0 nsec*

c) Jitter, stability:

*OK*

d) Prime power:

*+15 ± 0.5 V*

*275 mA*

e) PRF:

*0 to 40 kHz*

*100 VOLTS/DIV*

*2.0 nsec/DIV*

*40 kHz*