


PULSE GENERATOR
PERFORMANCE CHECK

Model: **AVO-2-17**

S.N.: **538**

Date: **June 30 81**



- a) Output signal amplitude:
0 TO +10 Amps
- b) Pulse width:
0 TO 15 NSEC
- c) Rise time:
OK
- d) Fall time:
OK
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 V

250 mA (max)

MID RANGE OUTPUT.

**32 VOLTS ACROSS
5 OHM.**

5 NSEC/DIV.

20 KHz.