


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVM-3-P

S.N.: 543

Date: JULY 15 81



10 VOLTS/DIV  
2.0 NSEC/DIV  
150 KHZ

- a) Output signal amplitude:  
0 TO 15 VOLTS
- b) Pulse width:  
0 TO 15 NSEC
- c) Rise time:  
≤ 150 PSEC
- d) Fall time:  
≤ 600 PSEC.
- e) PRF:\*  
0 TO 5 MHZ
- f) Jitter, stability:  
OK
- g) Prime power:  
+24 ± 1.0 V  
450 mA (MAX)

*[Handwritten signature]*

(\* S.N. 543 MODIFIED  
TO GIVE PRF MAX 5 MHZ  
& PW MAX 200 NSEC.)