


PULSE GENERATOR  
PERFORMANCE CHECK


Model: **AUP-AV-1-05-M**

S.N.: **548**

Date: **JUNE 30 81**



**MAIN OUT**  
**10 VOLTS/DIV**  
**1.0 NSEC/DIV**  
**400 KHz**



**M OUT**  
**1.0 VOLT/DIV**  
**1.0 NSEC/DIV**  
**400 KHz.**

- a) Output signal amplitude:  
**0.70 + 10 V**
- b) Pulse width:  
**0.2 TO 4.0 NSEC**
- c) Rise time:  
**≤ 100 PSEC**
- d) Fall time:  
**≤ 135 PSEC**
- e) PRF:  
**0 TO 1.0 MHz**
- f) Jitter, stability:  
**OK**
- g) Prime power:

**+15 ± 1.0 V**

**300 mA (MAX)**