

IMPULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S

S.N.: 557

Date: 1464 31 81

a) Output signal amplitude:

$\geq +15 \text{ V}$

b) Pulse width:

$\leq 500 \text{ psec}$


c) Jitter, stability:

OK.

d) Prime power:

$+15 \pm 1.0 \text{ V}$

150 mA (max)



500 psec/div

10 volts/div

400 kHz