


IMPULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-2-A-P*

S.N.: *560*

Date: *July 20 81*



a) Output signal amplitude:

0 to +120V

b) Pulse width:

≤ 2 nsec

c) Jitter, stability:

OK

d) Prime power:

+15 ± 1.0 VOLT.

225 mA (max)

*100 V / DIV
2.0 nsec / DIV
100 KHz.*